

SUNA1917

Version: V2

Operator:

Reference Update Time: 03-Jul-2023 23:26:31

Comment: Summary Report created automatically by the UCI Diagnostic Data Collection tool.

Ancillary	Value	Setting	Value
Firmware Revision	2.5.1	SerialBaudRate	57600
Total Disk Space (byte)	3954802688	OperationMode	Continuous
Free Disk Space (byte)	3677061120	FixedDuration	60
Used Lamp Time (sec)	329026	FrameTransfer	FULL_ASCII
Clock Time	16 Sep 2024 03:41:25 UTC	FrameLogging	FULL_ASCII
USB Enabled	true	OperationControl	Samples
PCB Supervisor	true	CycleLightFrames	30
Analog Board	true	CycleDarkFrames	5
SDI12 Board	true	CycleLightDuration	58
Relay Board	false	CycleDarkDuration	2
Super Capacitors	true	Baseline	Linear
Scheduling	true	Adaptive Integration	On
Data Logging	true	Concentrations Fitted	N03_SW_TSW
APF Interface	false	Concentrations Available	3
Integrated Wiper	true	Bromide Trace	false
External Power Port	false	IntegPeriod (ms)	325
Lamp Switch Off Temperature	35	WavelenMin (nm)	217.0
		WavelenMax (nm)	240.0
		NO3DacMin (μM)	-5.0
		NO3DacMax (μM)	100.0
		SpectralInLightAvg	1
		SpectralInDarkAvg	1
		PeriodicInterval	1 hr
		PeriodicOffset	0
		PeriodicDuration	5
		PeriodicSamples	15
		LogFileType	By Sample Event
		LogFileMaxSz	5

SUNA 1917

Version: V2

Configuration

SENSTYPE SUNA
SENSVERS V2
SERIALNO 1917
THEBRAND SeabirdCoastal
PATHLGTH 10mm
INTWIPER Available
EXTPPORT Missing
SUPRCAPS Available
PWRSVISR Available
USBSWTCH Available
RELAYBRD Missing
SDI12BRD Available
ANALGBRD Available
INTDATLG Available
APFIFACE Missing
SCHDLING Available
OWIRETLP 1054bfb603080023
OWIRETSP 10df22b1030800ec
OWIRETHS 108627c10308002e
ZSPEC_SN 141174
FIBERLSN L21.E01.212663
CUSTOMID None
STUPSTUS Done
BAUDRATE 57600
MSGLEVEL Info
MSGFSIZE 2
DATFSIZE 5
OUTFRITYP Full_ASCII
LOGFRITYP Full_ASCII
OUTDRKFR Output
LOGDRKFR Output
LOGFTYPE Acquisition
AFILEDUR 60
ACQCOUNT 11
CNTCOUNT 0
DCMINNO3 -5.000
DCMAXNO3 100.000
WDAT_LOW 217.00
WDAT_HGH 250.00
SDI12ADD 48
DATAMODE Real
OPERMODE Continuous
OPERCTRL Samples
EXDEVITYP None
EXDEVPRE 15
EXDEVRUN Off
EXDVIVAL 0

Configuration

COUNTDWN 3
FIXDDURA 60
PERDIVAL 1h
PERDOFFS 0
PERDDURA 5
PERDSMPL 15
POLLTOUT 15
APFATOFF 10.0000
STBLTIME 5
SKPSLEEP Off
LAMPTOFF 35
SPINTPER 325
DRKAVERS 1
LGTAVERS 1
DRKSMPLS 5
LGTSMPLS 30
DRKDURAT 2
LGTDURAT 58
TEMPCOMP Off
SALINFIT On
BRMTRACE Off
BL_ORDER 1
FITCONCS 3
DRKCORMT SpecAverage
A_CUTOFF 1.3000
INTPRADJ On
INTPRFAC 1
INTADSTP 5
INTADMAX 20
WFIT_LOW 217.00
WFIT_HGH 240.00
LAMPTIME 329026